

NEIDL Agent Incident Reporting Summary April 2015 - June 2015, Q2

Date of Incident	Incident Type / Agent Involved	BSL	Transmissible Person to Person	Description	Reportable Incident	Reportable of Clinical Illness	Comments / Corrective Actions
6/1/15	Splash to arm	Mechanical Room	No	<p>ROHP received a call from a scientist that a staff person working in the mechanical room sustained a couple of drops of liquid on to his arm from the ceiling above. The liquid was melting condensate from a freezer that contained sealed tubes of C. Diff specimens in a secondary box inside the freezer. The staff member was evaluated in ROHP.</p>	No	No	<p>All equipment should be properly maintained and in good working order with descriptive inventories and contact information available. Staff have been retrained to notify the control center and EHS in case of any emergency or other situation requiring assistance.</p>
6/15/15	Chemical Exposure	1st Floor	No	<p>A staff member reported that he picked up a sealed shipping box containing approximately 6 pint size containers of 35% H2O2 to bring up to store in a corrosive cabinet when he felt something wet on his left hand and right lower mid abdomen. He reports he got approximately 5ml of liquid on himself despite the containers having screw tops and being sealed in another plastic bag. He immediately washed with soap and water, changed out of his shirt and was evaluated and treated in ROHP. This staff member reported he has some white discoloration on his left thumb and middle fingers otherwise no complaints.</p>	No	No	<p>Upon receipt, shipments will be inspected for signs of leakage.</p>